



Report Number: ZUL7VVF

221 East Lincoln Avenue, Fort Collins, CO 80524 USA
1-970-498-1500, 1-800-446-7488, FAX: 1-970-498-1598
Visit us at www.in-situ.com

Instrument Details:

Instrument Model: Level TROLL 500
Full Scale Pressure Range: 5 PSI / 3.5 m / 11.5 ft / vented
Serial Number: 1160019
Hardware Version: 5
Firmware Version: 3.07

Calibration Details:

Calibration Result: PASS
Calibration Date: 2024-09-27 18:49:55 (UTC)
Nominal Range of Applied Temperature: -5 C to +50 C
Temperature Accuracy Specification: +/- 0.1 C From -5 C to +50 C
Nominal Range of Applied Pressure: 0 PSI to 5 PSI
Pressure Accuracy Specification: +/- 0.1 %FS from -5 C to +50 C, +/- 0.05 %FS at +15 C

Post-Calibration Check:

Parameter	Applied	Reported	Deviation	Unit
Pressure	5.0000	5.0002	0.0002	PSI
Pressure	2.0000	2.0000	0.0000	PSI
Pressure	-0.0001	-0.0005	-0.0004	PSI
Temperature	40.0756	40.0721	-0.0035	C

Calibration Procedures and Equipment Used:

Automated calibration procedures used.
Manu Agilent Model 34970A SerialNo MY59002551
Manu Mensor Model CPC6050 SerialNo 41001DL9
Manu Fluke Model 1504 SerialNo C17499
Manu Instrulab Model 406X-031-01 SerialNo 1-31079
Manu Agilent Model 53131A-010 SerialNo MY47002282
Manu MENSOR Model 600 SerialNo 620946

Notes:

- Standards used in this calibration are traceable to the National Institute of Standards and Technology.
- This calibration report shall not be reproduced, except in full, without the written approval of In-Situ, Inc.
- A calibration interval of 12 to 18 months is recommended.



Adukis S.

Performed By: AT

Report generated 10/4/2024 7:22:44 PM

Copyright © 2005-2024 In-Situ, Inc.

Page 1 of 1