



## Calibration

Report Number: 202312180330

221 East Lincoln Avenue, Fort Collins, CO 8

1-970-498-1500, 1-800-446-7488, FAX: 1-970

Visit us at [www.in-situ.com](http://www.in-situ.com)

### Instrument Details:

Instrument Model:

Full Scale Pressure Range:

Serial Number:

Level TROLL 500

15 PSI vented

890927

06-05-202

L5-G3

### Calibration Details:

Calibration Result:

Calibration Date:

Nominal Range of Applied Temperature:

Temperature Accuracy Specification:

Nominal Range of Applied Pressure:

Pressure Accuracy Specification:

PASS

2023-12-18 03:30:26 (UTC)

-5 C to +50 C

+/- 0.1 C From -5 C to +50 C

0.0 PSI to 15.0 PSI

+/- 0.1 %FS from -5 C to +50 C, +/- 0.05 %FS at +15 C

### Post-Calibration Check:

Parameter	Applied	Reported	Deviation
Pressure	15.0000	15.0001	0.0008
Pressure	6.3000	6.2990	-0.0066
Pressure	0.0003	-0.0003	-0.0040
Temperature	25.1560	25.1611	0.0051

### Calibration Procedures and Equipment Used:

Automated calibration procedures used.

Manu Agilent Model 34980A SerialNo MY44014053

Manu Instrulab Model 4312A-15 SerialNo 41039

Manu Instrulab Model 832-151-01 SerialNo 12157

Manu Mensor Model CPC6000 SerialNo 410008J4

### Notes:

1. Standards used in this calibration are traceable to the National Institute of Standards and Technology
2. This calibration report shall not be reproduced, except in full, without the written approval of In-Situ, Inc.
3. A calibration interval of 12 to 18 months is recommended.

Performed By: FM

Report generated: 2023-12-20 11:40:59 UTC

Copyright © 2005-2023 In-Situ, Inc.

P