



Calibration Report

Report Number: 2023051400658-868645
221 East Lincoln Avenue, Fort Collins, CO 80524 USA
1-970-498-1500, 1-800-446-7488, FAX: 1-970-498-1598
Visit us at www.in-situ.com

Instrument Details:

Instrument Model:
Full Scale Pressure Range:
Serial Number:

Level TROLL 500
5 PSI vented
868645

L7-G2

16-08-23

Calibration Details:

Calibration Result:
Calibration Date:
Nominal Range of Applied Temperature:
Temperature Accuracy Specification:
Nominal Range of Applied Pressure:
Pressure Accuracy Specification:

PASS
2023-05-14 00:06:58 (UTC)
-5 C to +50 C
+/- 0.1 C From -5 C to +50 C
0.0 PSI to 5.0 PSI
+/- 0.1 %FS from -5 C to +50 C, +/- 0.05 %FS at +15 C

Post-Calibration Check:

Parameter	Applied	Reported	Deviation
Pressure	5.0030	4.9985	-0.00293
Pressure	2.1030	2.0985	-0.00299
Pressure	0.0032	-0.0013	-0.0302
Temperature	24.7150	24.7214	0.0064

Calibration Procedures and Equipment Used:

Automated calibration procedures used:
Manu Agilent Model 34380A SerialNo MY44C02373
Manu Instrulab Model 4312A-15 SerialNo 41016
Manu Instrulab Model 832-151-01 SerialNo 12086
Manu Vensor Model CPC6000 SerialNo 41000834

Notes:

- Standards used in this calibration are traceable to the National Institute of Standards and Technology.
- This calibration report shall not be reproduced, except in full, without the written approval of In-Situ, Inc.
- A calibration interval of 12 to 18 months is recommended.

Performed By: FM

Report generated: 2023-05-15 11:53:16 UTC

Copyright © 2005-2023 In-Situ, Inc.

Page 1 of 1