



Calibration Report

Report Number: 20230919015859-890514

221 East Lincoln Avenue, Fort Collins, CO 80524 USA
1-970-498-1500, 1-800-446-7488, FAX: 1-970-498-1598

Visit us at www.in-situ.com

Instrument Details:

Instrument Model: Level TROLL 500
Full Scale Pressure Range: 15 PSI vented
Serial Number: 890514

Calibration Details:

Calibration Result: PASS
Calibration Date: 2023-09-19 01:58:59 (UTC)
Nominal Range of Applied Temperature: -5 C to +50 C
Temperature Accuracy Specification: +/- 0.1 C From -5 C to +50 C
Nominal Range of Applied Pressure: 0.0 PSI to 15.0 PSI
Pressure Accuracy Specification: +/- 0.1 %FS from -5 C to +50 C, +/- 0.05 %FS at +15 C

Pte SAN WLS
15.02.2024

Post-Calibration Check:

Parameter	Applied	Reported	Deviation
Pressure	15.0000	14.9991	-0.0062
Pressure	6.3000	6.2992	-0.0051
Pressure	0.0005	-0.0011	-0.0107
Temperature	25.1610	25.1651	0.0041

Calibration Procedures and Equipment Used:

Automated calibration procedures used.
Manu Agilent Model 34980A SerialNo MY44014053
Manu Instrulab Model 4312A-15 SerialNo 41039
Manu Instrulab Model 832-151-01 SerialNo 12157
Manu Mensor Model CPC6000 SerialNo 410008J4

Notes:

- Standards used in this calibration are traceable to the National Institute of Standards and Technology.
- This calibration report shall not be reproduced, except in full, without the written approval of In-Situ, Inc.
- A calibration interval of 12 to 18 months is recommended.

Performed By: FM

Report generated: 2023-09-19 15:03:30 UTC