

Instrument Details:

Instrument Model: **Level TROLL 500**
Full Scale Pressure Range: **5 PSI vented**
Serial Number: **937852**

Calibration Details:

Calibration Result: **PASS**
Calibration Date: 2024-10-18 20:9:23 (UTC)
Nominal Range of Applied Temperature: -5 C to +50 C
Temperature Accuracy Specification: +/- 0.1 C From -5 C to +50 C
Nominal Range of Applied Pressure: 0.0 PSI to 5.0 PSI
Pressure Accuracy Specification: +/- 0.1 %FS from -5 C to +50 C, +/- 0.05 %FS at +15 C

Post-Calibration Check:

Parameter	Applied	Reported	Deviation
Pressure	5.0000	4.9993	-0.0138
Pressure	2.1000	2.0987	-0.0250
Pressure	0.0002	-0.0008	-0.0193
Temperature	25.0680	25.0758	0.0078

Calibration Procedures and Equipment Used:

Automated calibration procedures used.
Manu Agilent Model 34980A SerialNo MY44000223
Manu Instrulab Model 4312A-15 SerialNo 41016
Manu Instrulab Model 832-151-01 SerialNo 12072
Manu Mensor Model CPC6000 SerialNo 410009W9

Notes:

- Standards used in this calibration are traceable to the National Institute of Standards and Technology.
- This calibration report shall not be reproduced, except in full, without the written approval of In-Situ, Inc.
- A calibration interval of 12 to 18 months is recommended.

Performed By: NA