



Calibration Report

Report Number: 20240925-1160257

221 East Lincoln Avenue, Fort Collins, CO 80524 USA

1-970-498-1500, 1-800-446-7488, FAX: 1-970-498-1598

Visit us at www.in-situ.com

Instrument Details:

Instrument Model: Level TROLL 500
Full Scale Pressure Range: 5 PSI / 3.5 m / 11.5 ft / vented
Serial Number: 1160257
Hardware Version: 5
Firmware Version: 3.07

08-02-2025

~~11-06-2024~~

L1-64 mg.

Calibration Details:

Calibration Result: PASS
Calibration Date: 2024-09-25 09:08:53 (UTC)
Nominal Range of Applied Temperature: -5 C to +50 C
Temperature Accuracy Specification: +/- 0.1 C From -5 C to +50 C
Nominal Range of Applied Pressure: 0 PSI to 5 PSI
Pressure Accuracy Specification: +/- 0.1 %FS from -5 C to +50 C, +/- 0.05 %FS at +15 C

Post-Calibration Check:

Parameter	Applied	Reported	Deviation	Unit
Pressure	5.0000	5.0004	0.0004	PSI
Pressure	2.0000	2.0004	0.0004	PSI
Pressure	0.0003	0.0004	0.0002	PSI
Temperature	39.8590	39.8560	-0.0030	C

Calibration Procedures and Equipment Used:

Automated calibration procedures used.

Manu Agilent Model 34970A SerialNo MY44000742

Manu Mensor Model CPC6000 SerialNo 4100187q

Manu Instrulab Model 3312A-14-15-24 SerialNo 31127

Manu Instrulab Model 406X-031-05 SerialNo 21663-21

Manu Agilent Model 53131A-010 SerialNo MY47002282

Manu MENSOR Model 600 SerialNo 620946

Notes:

- Standards used in this calibration are traceable to the National Institute of Standards and Technology.
- This calibration report shall not be reproduced, except in full, without the written approval of In-Situ, Inc.
- A calibration interval of 12 to 18 months is recommended.



Adukis S.

Performed By: JD

Report generated 10/4/2024 6:05:27 PM

Copyright © 2005-2024 In-Situ, Inc.

Page 1 of 1