

**Instrument Details:**

Instrument Model: **Level TROLL 500**  
Full Scale Pressure Range: **5 PSI / 3.5 m / 11.5 ft / vented**  
Serial Number: **1174145**  
Hardware Version: **5**  
Firmware Version: **3.07**

L14-4

03/03/25.

**Calibration Details:**

Calibration Result: **PASS**  
Calibration Date: **2024-11-18 20:43:43 (UTC)**  
Nominal Range of Applied Temperature: **-5 C to +50 C**  
Temperature Accuracy Specification: **+/- 0.1 C From -5 C to +50 C**  
Nominal Range of Applied Pressure: **0 PSI to 5 PSI**  
Pressure Accuracy Specification: **+/- 0.1 %FS from -5 C to +50 C, +/- 0.05 %FS at +15 C**

**Post-Calibration Check:**

Parameter	Applied	Reported	Deviation	Unit
Pressure	5.0000	4.9997	-0.0004	PSI
Pressure	2.0000	1.9998	-0.0002	PSI
Pressure	0.0002	0.0000	-0.0002	PSI
Temperature	39.8690	39.8631	-0.0059	C

**Calibration Procedures and Equipment Used:**

Automated calibration procedures used.

Manu Agilent Model 34970A SerialNo MY44000742

Manu Mensor Model CPC6000 SerialNo 4100187q

Manu Instrulab Model 3312A-14-15-24 SerialNo 31127

Manu Instrulab Model 406X-031-05 SerialNo 21663-21

Manu Agilent Model 53131A-010 SerialNo MY47002282

Manu MENSOR Model 600 SerialNo 620946

**Notes:**

- Standards used in this calibration are traceable to the National Institute of Standards and Technology.
- This calibration report shall not be reproduced, except in full, without the written approval of In-Situ, Inc.
- A calibration interval of 12 to 18 months is recommended.



Adukis S.

Performed By: AT