



## Calibration Report

Report Number: 20241104-1170734

221 East Lincoln Avenue, Fort Collins, CO 80524 USA

1-970-498-1500, 1-800-446-7488, FAX: 1-970-498-1598

Visit us at [www.in-situ.com](http://www.in-situ.com)

### Instrument Details:

Instrument Model: Level TROLL 500  
Full Scale Pressure Range: 5 PSI / 3.5 m / 11.5 ft / vented  
Serial Number: 1170734  
Hardware Version: 5  
Firmware Version: 3.07

P1-4  
22-02-25

### Calibration Details:

Calibration Result: PASS  
Calibration Date: 2024-11-04 15:08:17 (UTC)  
Nominal Range of Applied Temperature: -5 C to +50 C  
Temperature Accuracy Specification: +/- 0.1 C From -5 C to +50 C  
Nominal Range of Applied Pressure: 0 PSI to 5 PSI  
Pressure Accuracy Specification: +/- 0.1 %FS from -5 C to +50 C, +/- 0.05 %FS at +15 C

### Post-Calibration Check:

Parameter	Applied	Reported	Deviation	Unit
Pressure	5.0000	5.0002	0.0002	PSI
Pressure	2.0000	2.0002	0.0002	PSI
Pressure	0.0000	0.0005	0.0005	PSI
Temperature	39.8883	39.8780	-0.0103	C

### Calibration Procedures and Equipment Used:

Automated calibration procedures used.  
Manu Agilent Model 34970A SerialNo MY41027658  
Manu Mensor Model CPC6050 SerialNo 41001ABW  
Manu Fluke Model 1504 SerialNo C14479  
Manu Instrulab Model 406X-031-05 SerialNo 21332-1  
Manu Agilent Model 53131A-010 SerialNo MY47002282  
Manu MENSOR Model 600 SerialNo 620946

### Notes:

- Standards used in this calibration are traceable to the National Institute of Standards and Technology.
- This calibration report shall not be reproduced, except in full, without the written approval of In-Situ, Inc.
- A calibration interval of 12 to 18 months is recommended.



Aduki's S.

Performed By: TG

Report generated 12/19/2024 2:57:44 PM

Copyright © 2005-2024 In-Situ, Inc.

Page 1 of 1