

**Instrument Details:**

Instrument Model: **Level TROLL 500**  
Full Scale Pressure Range: **5 PSI / 3.5 m / 11.5 ft / vented**  
Serial Number: **1159993**  
Hardware Version: **5**  
Firmware Version: **3.07**

*L2-26.**10/12/24.***Calibration Details:**

Calibration Result: **PASS**  
Calibration Date: **2024-09-25 01:57:08 (UTC)**  
Nominal Range of Applied Temperature: **-5 C to +50 C**  
Temperature Accuracy Specification: **+/- 0.1 C From -5 C to +50 C**  
Nominal Range of Applied Pressure: **0 PSI to 5 PSI**  
Pressure Accuracy Specification: **+/- 0.1 %FS from -5 C to +50 C, +/- 0.05 %FS at +15 C**

**Post-Calibration Check:**

Parameter	Applied	Reported	Deviation	Unit
Pressure	5.0000	4.9997	-0.0003	PSI
Pressure	2.0000	1.9999	-0.0001	PSI
Pressure	0.0003	-0.0002	-0.0005	PSI
Temperature	39.8670	39.8631	-0.0039	C

**Calibration Procedures and Equipment Used:**

Automated calibration procedures used.  
Manu Agilent Model 34970A SerialNo MY44000742  
Manu Mensor Model CPC6000 SerialNo 4100187q  
Manu Instrulab Model 3312A-14-15-24 SerialNo 31127  
Manu Instrulab Model 406X-031-05 SerialNo 21663-21  
Manu Agilent Model 53131A-010 SerialNo MY47002282  
Manu MENSOR Model 600 SerialNo 620946

**Notes:**

- Standards used in this calibration are traceable to the National Institute of Standards and Technology.
- This calibration report shall not be reproduced, except in full, without the written approval of In-Situ, Inc.
- A calibration interval of 12 to 18 months is recommended.

*Adukis S.***Performed By:** DES