



## Calibration Report

Report Number: 20250210-1198152

221 East Lincoln Avenue, Fort Collins, CO 80524 USA  
1-970-498-1500, 1-800-446-7488, FAX: 1-970-498-1598

Visit us at [www.in-situ.com](http://www.in-situ.com)

### Instrument Details:

Instrument Model: **Level TROLL 500**  
Full Scale Pressure Range: **5 PSI / 3.5 m / 11.5 ft / vented**  
Serial Number: **1198152**  
Hardware Version: **5**  
Firmware Version: **3.07**

L9-2

06-06-25

### Calibration Details:

Calibration Result: **PASS**  
Calibration Date: **2025-02-10 16:02:02 (UTC)**  
Nominal Range of Applied Temperature: **-5 C to +50 C**  
Temperature Accuracy Specification: **+/- 0.1 C From -5 C to +50 C**  
Nominal Range of Applied Pressure: **0 PSI to 5 PSI**  
Pressure Accuracy Specification: **+/- 0.1 %FS from -5 C to +50 C, +/- 0.05 %FS at +15 C**

### Post-Calibration Check:

Parameter	Applied	Reported	Deviation	Unit
Pressure	5.0000	4.9995	-0.0005	PSI
Pressure	2.0000	2.0000	0.0000	PSI
Pressure	-0.0001	0.0003	0.0004	PSI
Temperature	40.0727	40.0667	-0.0060	C

### Calibration Procedures and Equipment Used:

Automated calibration procedures used.

Manu Agilent Model 34970A SerialNo MY59002551

Manu Mensor Model CPC6050 SerialNo 41001DL9

Manu Fluke Model 1504 SerialNo C49735

Manu Instrulab Model 406X-031-01 SerialNo 19528-1

Manu Agilent Model 53131A-010 SerialNo MY47002282

Manu MENSOR Model 600 SerialNo 620946

### Notes:

- Standards used in this calibration are traceable to the National Institute of Standards and Technology.
- This calibration report shall not be reproduced, except in full, without the written approval of In-Situ, Inc.
- A calibration interval of 12 to 18 months is recommended.



Adukis S.

Performed By: AT

Report generated 3/6/2025 8:19:42 PM

Copyright © 2005-2025 In-Situ, Inc.

Page 1 of 1