

Instrument Details:

Instrument Model:

Full Scale Pressure Range

Serial Number:

Level TROLL 500

5 PSI vented

685809

Calibration Details:

Calibration Result:

Calibration Date:

PASS

2023-09-16 01:37:44 (UTC)

Nominal Range of Applied Temperature:

-5 C to +50 C

Temperature Accuracy Specification:

+/- 0.1 C From -5 C to +50 C

Nominal Range of Applied Pressure:

0.0 PSI to 5.0 PSI

Pressure Accuracy Specification:

+/- 0.1 %FS from -5 C to +50 C, +/- 0.05 %FS at +15 C

Post-Calibration Check:

Parameter	Applied	Reported	Deviation
Pressure	5.0000	4.9994	-0.0123
Pressure	2.1000	2.0999	-0.0023
Pressure	-0.0001	-0.0002	-0.0004
Temperature	25.1630	25.1538	-0.0092

Calibration Procedures and Equipment Used:

Automated calibration procedures used.

Manu Agilent Model 34980A SerialNo MY44014053

Manu Instrulab Model 4312A-15 SerialNo 41039

Manu Instrulab Model 832-151-01 SerialNo 12157

Manu Mensor Model CPC6000 SerialNo 410008J4

Notes:

1. Standards used in this calibration are traceable to the National Institute of Standards and Technology.
2. This calibration report shall not be reproduced, except in full, without the written approval of In-Situ, Inc.
3. A calibration interval of 12 to 18 months is recommended.

Performed By: FM