

**Instrument Details:**

Instrument Model: Level TROLL 500  
Full Scale Pressure Range: 15 PSI vented  
Serial Number: 890510

L7-4  
14-11-23

**Calibration Details:**

Calibration Result: **PASS**  
Calibration Date: 2023-09-19 01:58:59 (UTC)  
Nominal Range of Applied Temperature: -5 C to +50 C  
Temperature Accuracy Specification: +/- 0.1 C From -5 C to +50 C  
Nominal Range of Applied Pressure: 0.0 PSI to 15.0 PSI  
Pressure Accuracy Specification: +/- 0.1 %FS from -5 C to +50 C, +/- 0.05 %FS at +15 C

**Post-Calibration Check:**

Parameter	Applied	Reported	Deviation
Pressure	15.0000	14.9991	-0.0063
Pressure	6.3000	6.2984	-0.0108
Pressure	0.0005	-0.0014	-0.0126
Temperature	25.1620	25.1665	0.0045

**Calibration Procedures and Equipment Used:**

Automated calibration procedures used.  
Manu Agilent Model 34980A SerialNo MY44014053  
Manu Instrulab Model 4312A-15 SerialNo 41039  
Manu Instrulab Model 832-151-01 SerialNo 12157  
Manu Mensor Model CPC6000 SerialNo 410008J4

**Notes:**

1. Standards used in this calibration are traceable to the National Institute of Standards and Technology.
2. This calibration report shall not be reproduced, except in full, without the written approval of In-Situ, Inc.
3. A calibration interval of 12 to 18 months is recommended.

**Performed By: FM**