

Instrument Details:

Instrument Model: Level TROLL 500
Full Scale Pressure Range: 15 PSI vented
Serial Number: 891313

L7-6

17-02-24

Calibration Details:

Calibration Result: **PASS**
Calibration Date: 2023-12-18 03:30:26 (UTC)
Nominal Range of Applied Temperature: -5 C to +50 C
Temperature Accuracy Specification: +/- 0.1 C From -5 C to +50 C
Nominal Range of Applied Pressure: 0.0 PSI to 15.0 PSI
Pressure Accuracy Specification: +/- 0.1 %FS from -5 C to +50 C, +/- 0.05 %FS at +15 C

Post-Calibration Check:

Parameter	Applied	Reported	Deviation
Pressure	15.0000	14.9977	-0.0155
Pressure	6.3000	6.2976	-0.0160
Pressure	0.0003	-0.0022	-0.0169
Temperature	25.1570	25.1647	0.0077

Calibration Procedures and Equipment Used:

Automated calibration procedures used.
Manu Agilent Model 34980A SerialNo MY44014053
Manu Instrulab Model 4312A-15 SerialNo 41039
Manu Instrulab Model 832-151-01 SerialNo 12157
Manu Mensor Model GPC6000 SerialNo 410008J4

Notes:

1. Standards used in this calibration are traceable to the National Institute of Standards and Technology.
2. This calibration report shall not be reproduced, except in full, without the written approval of In-Situ, Inc.
3. A calibration interval of 12 to 18 months is recommended.

Performed By: FM