



Calibration Report

Report Number: 20241207005235-593898

221 East Lincoln Avenue, Fort Collins, CO 80524 USA
1-970-498-1500, 1-800-446-7488, FAX: 1-970-498-1598

Visit us at www.in-situ.com

Instrument Details:

Instrument Model: Level TROLL 500
Full Scale Pressure Range: 5 PSI vented
Serial Number: 593898

L5-G3
PO201
25/05/25.

Calibration Details:

Calibration Result: PASS
Calibration Date: 2024-12-07 00:52:35 (UTC)
Nominal Range of Applied Temperature: -5 C to +50 C
Temperature Accuracy Specification: +/- 0.1 C From -5 C to +50 C
Nominal Range of Applied Pressure: 0.0 PSI to 5.0 PSI
Pressure Accuracy Specification: +/- 0.1 %FS from -5 C to +50 C, +/- 0.05 %FS at +15 C

Post-Calibration Check:

Parameter	Applied	Reported	Deviation
Pressure	5.0000	4.9990	-0.0200
Pressure	2.1000	2.0990	-0.0199
Pressure	0.0001	-0.0010	-0.0218
Temperature	25.1230	25.1310	0.0080

Calibration Procedures and Equipment Used:

Automated calibration procedures used.
Manu Agilent Model 34980A SerialNo MY44002373
Manu Instrulab Model 4312A-15 SerialNo 41039
Manu Instrulab Model 832-151-01 SerialNo 12084
Manu Mensor Model CPC6000 SerialNo 410008J4

Notes:

- Standards used in this calibration are traceable to the National Institute of Standards and Technology.
- This calibration report shall not be reproduced, except in full, without the written approval of In-Situ, Inc.
- A calibration interval of 12 to 18 months is recommended.

Performed By: FM

Report generated: 2025-01-21 10:49:29 UTC

Copyright © 2005-2025 In-Situ, Inc.

Page 1 of 1