

**Instrument Details:**Instrument Model  
Full Scale Pressure Range:  
Serial Number:Level TROLL 600  
5 PSI vented  
685824

L7- 62- 10-20 15-05-25

**Calibration Details:**Calibration Result:  
Calibration Date:  
Nominal Range of Applied Temperature:  
Temperature Accuracy Specification:  
Nominal Range of Applied Pressure:  
Pressure Accuracy Specification:**PASS**  
2025-03-21 22:15:54 (UTC)  
-5 C to +50 C  
+/- 0.1 C From -5 C to +50 C  
0.0 PSI to 5.0 PSI  
+/- 0.1 %FS from -5 C to +50 C, +/- 0.05 %FS at +15 C**Post-Calibration Check:**

Parameter	Applied	Reported	Deviation
Pressure	5.0000	5.0009	0.0168
Pressure	2.1000	2.1011	0.0222
Pressure	0.0003	0.0020	0.0345
Temperature	25.1240	25.1256	0.0016

**Calibration Procedures and Equipment Used:**Automated calibration procedures used.  
Manu Agilent Model 34980A SerialNo MY44014053  
Manu Instrulab Model 4312A-15 SerialNo 41039  
Manu Instrulab Model 832-151-01 SerialNo 12084  
Manu Mensor Model CPC6000 SerialNo 410008J4**Notes:**

- Standards used in this calibration are traceable to the National Institute of Standards and Technology.
- This calibration report shall not be reproduced, except in full, without the written approval of In-Situ, Inc.
- A calibration interval of 12 to 18 months is recommended.