

Instrument Details:

Instrument Model: **Level TROLL 500**
Full Scale Pressure Range: **5 PSI / 3.5 m / 11.5 ft / vented**
Serial Number: **1162477**
Hardware Version: **5**
Firmware Version: **3.07**

P 2-2
01/03/2025

Calibration Details:

Calibration Result: **PASS**
Calibration Date: **2024-10-07 15:54:05 (UTC)**
Nominal Range of Applied Temperature: **-5 C to +50 C**
Temperature Accuracy Specification: **+/- 0.1 C From -5 C to +50 C**
Nominal Range of Applied Pressure: **0 PSI to 5 PSI**
Pressure Accuracy Specification: **+/- 0.1 %FS from -5 C to +50 C, +/- 0.05 %FS at +15 C**

Post-Calibration Check:

Parameter	Applied	Reported	Deviation	Unit
Pressure	5.0003	5.0007	0.0005	PSI
Pressure	2.0000	2.0006	0.0006	PSI
Pressure	0.0006	0.0010	0.0004	PSI
Temperature	40.0578	40.0393	-0.0185	C

Calibration Procedures and Equipment Used:

Automated calibration procedures used.
Manu Agilent Model 34970A SerialNo MY41029005
Manu Mensor Model CPC6000 SerialNo 610379
Manu Fluke Model 1504 SerialNo C14476
Manu Instrulab Model 406 SerialNo 21663-1
Manu Agilent Model 53131A-010 SerialNo MY47002282
Manu MENSOR Model 600 SerialNo 620946

Notes:

- Standards used in this calibration are traceable to the National Institute of Standards and Technology.
- This calibration report shall not be reproduced, except in full, without the written approval of In-Situ, Inc.
- A calibration interval of 12 to 18 months is recommended.



Adukis S.

Performed By: TG