

**Instrument Details:**

Instrument Model: **Level TROLL 500**  
Full Scale Pressure Range: **5 PSI vented**  
Serial Number: **749163**

PULLAN 15-05-25

**Calibration Details:**

Calibration Result: **PASS**  
Calibration Date: 2025-03-23 00:6:19 (UTC)  
Nominal Range of Applied Temperature: -5 C to +50 C  
Temperature Accuracy Specification:  $\pm 0.1$  C From -5 C to +50 C  
Nominal Range of Applied Pressure: 0.0 PSI to 5.0 PSI  
Pressure Accuracy Specification:  $\pm 0.1$  %FS from -5 C to +50 C,  $\pm 0.05$  %FS at +15 C

**Post-Calibration Check:**

Parameter	Applied	Reported	Deviation
Pressure	5.0000	4.9991	-0.0175
Pressure	2.1000	2.0983	-0.0321
Pressure	0.0000	-0.0015	-0.0301
Temperature	25.1220	25.1231	0.0011

**Calibration Procedures and Equipment Used:**

Automated calibration procedures used  
Manu Agilent Model 34980A SerialNo MY44014053  
Manu Instrulab Model 4312A-15 SerialNo 41039  
Manu Instrulab Model 832-151-01 SerialNo 12084  
Manu Mensor Model CPC6000 SerialNo 410008J4

**Notes:**

- Standards used in this calibration are traceable to the National Institute of Standards and Technology
- This calibration report shall not be reproduced, except in full, without the written approval of In-Situ, Inc.
- A calibration interval of 12 to 18 months is recommended.