



Calibration Report

Report Number: 2025032300619-592964

221 East Lincoln Avenue, Fort Collins, CO 80524 USA
1-970-498-1500, 1-800-446-7488, FAX: 1-970-498-1598
Visit us at www.in-situ.com

Instrument Details:

Instrument Model:
Full Scale Pressure Range:
Serial Number:

Level TROLL 500
5 PSI vented
592964

P1-2
12-08-25

Calibration Details:

Calibration Result: **PASS**
Calibration Date: 2025-03-23 00:6:19 (UTC)
Nominal Range of Applied Temperature: -5 C to +50 C
Temperature Accuracy Specification: +/- 0.1 C From -5 C to +50 C
Nominal Range of Applied Pressure: 0.0 PSI to 5.0 PSI
Pressure Accuracy Specification: +/- 0.1 %FS from -5 C to +50 C, +/- 0.05 %FS at +15 C

Post-Calibration Check:

Parameter	Applied	Reported	Deviation
Pressure	5.0000	5.0007	0.0143
Pressure	2.1001	2.1010	0.0183
Pressure	0.0000	0.0009	0.0172
Temperature	25.1220	25.1303	0.0083

Calibration Procedures and Equipment Used:

Automated calibration procedures used.
Manu Agilent Model 34980A SerialNo MY44014053
Manu Instrulab Model 4312A-15 SerialNo 41039
Manu Instrulab Model 832-151-01 SerialNo 12084
Manu Mensor Model CPC6000 SerialNo 410008J4

Notes:

- Standards used in this calibration are traceable to the National Institute of Standards and Technology.
- This calibration report shall not be reproduced, except in full, without the written approval of In-Situ, Inc.
- A calibration interval of 12 to 18 months is recommended.

Performed By: FM

Report generated: 2025-03-24 16:04:04 UTC

Copyright © 2005-2025 In-Situ, Inc.

Page 1 of 1