Calibration Report

Report Number: 20260409204030-1027080
221 East Lincoln Avenue, Fort Collins, CO 80524 USA
1-970-498-1500, 1-800-446-7488, FAX: 1-970-498-1598
Visit us at www.in-situ.com

Instrument Details:

Instrument Model:	**Level TROLL 500**
Full Scale Pressure Range:	**5 PSI vented**
Serial Number:	**1027080**

Calibration Details:

Calibration Result:	**PASS**
Calibration Date:	2026-04-09 20:40:30 (UTC)
Nominal Range of Applied Temperature:	-5 C to +50 C
Temperature Accuracy Specification:	+/- 0.1 C From -5 C to +50 C
Nominal Range of Applied Pressure:	0.0 PSI to 5.0 PSI
Pressure Accuracy Specification:	+/- 0.05 %FS from 0 C to +50 C

Post-Calibration Check:

Parameter	Applied	Reported	Deviation
Pressure	5.0000	5.0000	-0.0001
Pressure	2.1000	2.1001	0.0024
Pressure	0.0002	0.0004	0.0025
Temperature	25.0590	25.0617	0.0027

Calibration Procedures and Equipment Used:

Automated calibration procedures used.
Manu Agilent Model 34980A SerialNo MY65220020
Manu Instrulab Model 4312A-15 SerialNo 41036
Manu Instrulab Model 832-151-01 SerialNo 12086
Manu Mensor Model CPC6000 SerialNo 410008J4

Notes:

1. Standards used in this calibration are traceable to the National Institute of Standards and Technology.
2. This calibration report shall not be reproduced, except in full, without the written approval of In-Situ, Inc.
3. A calibration interval of 12 to 18 months is recommended.

Performed By: FM

Copyright © 2005-2026 In-Situ, Inc.